




Inogen

Inogen One G-3

Report No. INOG0018



Report Prepared By Northwest EMC Inc.

NORTHWEST EMC – (888) 364-2378 – www.nwemc.com

California – Minnesota – Oregon – New York – Washington



22975 NW Evergreen Parkway
Suite 400
Hillsboro, Oregon 97124

Certificate of Test

Last Date of Test: September 29, 2011

Inogen

Model: Inogen One G-3

Emissions

Test Description	Specification	Test Method	Pass/Fail
Radiated Emissions	RTCA DO-160F:2007, Category M	RTCA DO-160F:2007	Pass
Conducted Emissions	RTCA DO-160F:2007, Category M	RTCA DO-160F:2007	Pass

Deviations From Test Standards

None

Approved By:

Victor Ratinoff, Operations Manager

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government of the United States of America. Product compliance is the responsibility of the client, therefore the tests and equipment modes of operation represented in this report were agreed upon by the client, prior to testing. This Report may only be duplicated in its entirety. The results of this test pertain only to the sample(s) tested. The specific description is noted in each of the individual sections of the test report supporting this certificate of test.

Revision Number	Description	Date	Page Number
00	None		



BOGOTÁ: (601) 390 56 20 • CALI: (602) 485 06 08
• MEDELLÍN: (604) 607 12 62 • PEREIRA: (606) 340 04 07 • IBAGUÉ: (608) 277 12 82
• PASTO: (602) 738 23 83 Ext: 7 • MIAMI: +1 305 4598648

CEL.:  +57 315 783 7832

CUSTOMER SERVICE LINE: TOLL FREE # 1-800-591-3409

 oxigenopoc@gmail.com  www.oxigenopoc.com

